

Venera Factory

Contact@venerafactory.com

949-600-0064

Sample: 08-30-2023-37818

Sample Received:08/30/2023;

Report Created: 09/11/2023; Expires: 08/30/2024

MC

Plant , Flower - Cured



21.478 %

Total THC

0.135 %

Δ-9 THC

25.011 %

Total Cannabinoids

<LOQ %

Total CBD

Cannabinoids

(Testing Method:HPLC, CON-P-3000)

Date Tested: 08/30/2023

Complete

Analyte	LOD	LOQ	Mass	Mass	
	%	%	%	mg/g	
Δ-8-Tetrahydrocannabinol (Δ-8-THC)	0.0488	0.0732	ND	ND	
Δ-9-Tetrahydrocannabinol (Δ-9-THC)	0.0488	0.0732	0.135	1.346	
Δ-9-Tetrahydrocannabinolic Acid (THCA-A)	0.0488	0.0732	24.337	243.366	
Δ-9-Tetrahydrocannabinophorol (Δ-9-THCP)	0.0488	0.0732	ND	ND	
Δ-9-Tetrahydrocannabivarin (Δ-9-THCV)	0.0488	0.0732	ND	ND	
Δ-9-Tetrahydrocannabivarinic Acid (Δ-9-THCVA)	0.0488	0.0732	<LOQ	<LOQ	
R-Δ-10-Tetrahydrocannabinol (R-Δ-10-THC)	0.0488	0.0732	ND	ND	
S-Δ-10-Tetrahydrocannabinol (S-Δ-10-THC)	0.0488	0.0732	ND	ND	
9R-Hexahydrocannabinol (9R-HHC)	0.0488	0.0732	ND	ND	
9S-Hexahydrocannabinol (9S-HHC)	0.0488	0.0732	ND	ND	
Tetrahydrocannabinol Acetate (THCO)	0.0488	0.0732	ND	ND	
Cannabidivarin (CBDV)	0.0488	0.0732	ND	ND	
Cannabidivarinic Acid (CBDVA)	0.0488	0.0732	ND	ND	
Cannabidiol (CBD)	0.0488	0.0732	ND	ND	
Cannabidiolic Acid (CBDA)	0.0322	0.0732	<LOQ	<LOQ	
Cannabigerol (CBG)	0.0488	0.0732	<LOQ	<LOQ	
Cannabigerolic Acid (CBGA)	0.0488	0.0732	0.396	3.961	
Cannabinol (CBN)	0.0488	0.0732	ND	ND	
Cannabinolic Acid (CBNA)	0.0488	0.0732	ND	ND	
Cannabichromene (CBC)	0.0488	0.0732	ND	ND	
Cannabichromenic Acid (CBCA)	0.0488	0.0732	0.143	1.434	
Total			25.011	250.107	

Total THC = THCa * 0.877 + Δ9-THC; Total CBD = CBDA * 0.877 + CBD; LOQ = Limit of Quantitation; ND = Not Detected.

Total THC Measurement of Uncertainty: ± 0.050%
 Total CBD Measurement of Uncertainty: ± 2.000%
 THCO potency analysis does not designate quantitative specificity of Δ-8-THCO and Δ-9-THCO isomers

Amended report issued to reflect change in sample identification.



New Bloom Labs
 6121 Heritage Park Drive, A500
 Chattanooga, TN 37416
 (844) 837-8223
 TN DEA#: RN0563975
 ANAB Testing Laboratory (AT-2868): ISO/IEC
 17025:2017

Natalie Siracusa
 Natalie Siracusa
 Laboratory Director

Powered by
 reLIMS
 info@relims.com